

Cold Van Operating Specification:

The van is intended to have accurate and stable temperature control for the processing of scientific samples. The cold area of the van shall be thermostatically controlled between 0°C (32°F) and 10°C (50°F). Once the temperature is set, it shall be capable of maintaining an internal temperature accuracy of +/- 1°C with a maximum outside temperature of 38°C (100°F), and a minimum outside temperature of 0°C (32°F).